DC 0089

Page 1 of 2

<u>CoperSurgical</u>

Declaration of Conformity

Application of Council Directive(s):

Low Voltage Directive 2006/95/EC

Machinery Directive 2006/42/EC

Electromagnetic Compatibility Directive 2004/108/EC

Manufacturer's Name:

CooperSurgical, Inc.

Manufacturer's Address:

95 Corporate Drive, Trumbull, CT 06611

Standards to which Conformity is Declared:

IEC 61010-2-010:2003, IEC 61010-1:2010 (3rd Edition)

Type of Equipment:

Heated Trolley and Warming Plates Family

Model Numbers:

HT37, WP37 300, WP37 500

Refer to Page 2 for device descriptions to the corresponding

part numbers.

Date of First Issue:

January 8, 2016

We hereby declare that the equipment specified above conforms to the above Directive(s) and Standard(s).

Rev. A: First Issue

(Signature)

(Date)

J. Chris Jensen

(Full Name)

Senior Director, Regulatory Affairs & Compliance (Position)

Declaration of Conformity

Heated Trolley and Warming Plates Family

Part Numbers (article number)	Device Description
HT37	Heated Trolley
(9002120147)	(Working area: 700mm x 600mm, Electrical Specs: 240V, 50-60Hz, Max: 6.1A)
WP37 300	Warming Plate
(9001701900)	(Working area: 300mm x 200mm, Electrical Specs: 240V, 50-60Hz, Max: 0.5A)
WP37 500	Warming Plate
(9001701910)	(Working area: 500mm x 400mm, Electrical Specs: 240V, 50-60Hz, Max: 1.7A)